

diclofenac

Reference Tilquin B. et. al., Chromatographia 36 (1993)120

Column Luna C18(2)


Mobile Phase 0.02 M Sodium Acetate Buffer (pH 5.0) / ACN (67:33)

Detection 229 nm

The phenyl column should give very good separation on this aromatic compound.

Gemini C6-Phenyl.

mobile phase of ACN/MeOH/NaAcetate Buffer (25:5:70).



primidone (application 10905, 14223)

